



**KEMENTERIAN KESEHATAN REPUBLIK INDONESIA
DIREKTORAT JENDERAL KEFARMASIAN DAN ALAT KESEHATAN**

Jalan H.R. Rasuna Said Blok X-5 Kavling 4 - 9 Jakarta 12950
Telepon : (021) 5201590 Pesawat 2029, 8011
Faksimile : (021) 52964838 Kotak Pos : 203



Berdasarkan Peraturan Menteri Kesehatan R.I Nomor 62 Tahun 2017 Tentang Izin Edar Alat Kesehatan, Alat Kesehatan Diagnostik In Vitro Dan Perbekalan Kesehatan Rumah Tangga dengan ini diberikan persetujuan untuk diedarkan dengan :

NOMOR IZIN EDAR

ALAT KESEHATAN

KEMENKES RI AKL 20805911761

Nama Dagang / Merek : **AXIMED Feeding Tube**
Kelompok / Kelas Resiko : Non Elektromedik Steril / B
Kategori Produk : Peralatan Gastroenterologi-Urologi
Sub Kategori : Peralatan Gastroenterologi-Urologi Terapetik
Jenis Produk : Gastrointestinal tube and accessories.
Tipe / Ukuran : Terlampir
Kemasan : Kantong Plastik
Nama Produsen / Pabrikan : WELL LEAD MEDICAL CO.,LTD, China
Melalui NINGBO GREETMED MEDICAL INSTRUMENTS CO., LTD,
China
Nama Pendaftar : PT. BIO AXION HEALTHINDO, DKI Jakarta
Atas dasar lisensi dari : -

Ketentuan

1. Persetujuan izin edar berlaku sampai dengan 29 Oktober 2024.
2. Wajib menyampaikan laporan berkala dan laporan jika ada kejadian yang tidak diinginkan akibat penggunaan Alat Kesehatan tersebut di atas sesuai ketentuan berlaku.
3. Izin edar ini merupakan persetujuan perubahan penandaan dari Nomor Izin Edar Alat Kesehatan KEMENKES RI AKL 20805911761 tanggal 29 Oktober 2021. Dengan demikian izin edar sebelumnya dinyatakan tidak berlaku.
4. Apabila dikemudian hari ada pihak lain yang berhak atas merek dan/atau keagenan produk tersebut, pendaftar bersedia mengembalikan izin edar.
5. Penandaan dan informasi produk yang terlampir merupakan bagian yang tidak terpisahkan dari persetujuan izin edar ini.
6. Izin Edar ini merupakan persetujuan produk OEM Impor.
7. Apabila di kemudian hari terdapat kekeliruan, maka persetujuan izin edar ini akan ditinjau kembali.

Jakarta, 25 November 2021

	Ditandatangani Secara Elektronik Oleh :	
	a.n	Direktur Jenderal
		Direktur Penilaian Alat Kesehatan dan PKRT
		Ir. Sodikin Sadek, M.Kes NIP.19621203 198603 1 004



Catatan:

- UU ITE No 11 Tahun 2007 Pasal 5 ayat 1
- Informasi Elektronik dan/atau Dokumen Elektronik dan/atau hasil cetaknya merupakan alat bukti hukum yang sah.
- Dokumen ini telah ditandatangani secara elektronik menggunakan sertifikat elektronik yang diterbitkan BSR.E.



KEMENTERIAN KESEHATAN REPUBLIK INDONESIA
DIREKTORAT JENDERAL KEFARMASIAN DAN ALAT KESEHATAN

Jalan H.R. Rasuna Said Blok X-5 Kavling 4 - 9 Jakarta 12950
Telepon : (021) 5201590 Pesawat 2029, 8011
Faksimile : (021) 52964838 Kotak Pos : 203



LAMPIRAN

NOMOR IZIN EDAR
ALAT KESEHATAN
KEMENKES RI AKL 20805911761

No.	Deskripsi	Tipe / Kode
1	AXIMED ® FEEDING TUBE, 40cm, no. 4	
2	AXIMED ® FEEDING TUBE, 40cm, no. 5	
3	AXIMED ® FEEDING TUBE, 40cm, no. 6	
4	AXIMED ® FEEDING TUBE, 40cm, no. 8	
5	AXIMED ® FEEDING TUBE, 40cm, no. 10	
6	AXIMED ® FEEDING TUBE, 40cm, no. 12	
7	AXIMED ® FEEDING TUBE, 40cm, no. 14	
8	AXIMED ® FEEDING TUBE, 40cm, no. 16	
9	AXIMED ® FEEDING TUBE, 40cm, no. 18	
10	AXIMED ® FEEDING TUBE, 100cm, no. 4	
11	AXIMED ® FEEDING TUBE, 100cm, no. 5	
12	AXIMED ® FEEDING TUBE, 100cm, no. 6	
13	AXIMED ® FEEDING TUBE, 100cm, no. 8	
14	AXIMED ® FEEDING TUBE, 100cm, no. 10	
15	AXIMED ® FEEDING TUBE, 100cm, no. 12	
16	AXIMED ® FEEDING TUBE, 100cm, no. 14	
17	AXIMED ® FEEDING TUBE, 100cm, no. 16	
18	AXIMED ® FEEDING TUBE, 100cm, no. 18	

Dengan ketentuan bahwa izin edar tersebut hanya berlaku untuk deskripsi dan tipe / kode produk yang tercantum dalam lampiran ini

Jakarta, 25 November 2021



KEMENTERIAN
KESEHATAN
REPUBLIK
INDONESIA

Ditandatangani Secara Elektronik Oleh :
a.n Direktur Jenderal
Direktur Penilaian Alat Kesehatan dan PKRT

Ir. Sodikin Sadek, M.Kes
NIP.19621203 198603 1 004



Catatan:
- UU ITE No 11 Tahun 2007 Pasal 5 ayat 1
Informasi Elektronik dan/atau Dokumen Elektronik dan/atau hasil cetaknya merupakan alat bukti hukum yang sah.
- Dokumen ini telah ditandatangani secara elektronik menggunakan sertifikat elektronik yang diterbitkan BSR.E.

25*13*6cm

CAUTION :
• Sterile, single use only
• Do not use if package is damage or opened

FEATURES :
• Made from non-toxic, non-irritant soft PVC
• Two-eyed tube used for nasogastric feeding of neonates and infants
• Transparent tube for easy observation of fluids
• Sterile, single use only
• Do not use if package is damage or opened

product of :
AXIONHealthindo

REF :
OD :
type :
size :

LATEXfree CE0123
STERILE EO

Reg No. HKMKS RI AKL 2005011761

PT. AXION HEALTHINDO
16 P.1 NO.1 Building 1, No.00 Changsan Road,
H. Taid Zone, 31562 Ningbo, Zhejiang Province, China

Disprodksi oleh :
C-4 Jatin Industrial Estate, Hualing Pampu,
Guangzhou, China

Model :
MINMED GREETMED MEDICAL INSTRUMENTS CO., LTD.
16 P.1 NO.1 Building 1, No.00 Changsan Road,
H. Taid Zone, 31562 Ningbo, Zhejiang Province, China

Disprodksi oleh :
C-4 Jatin Industrial Estate, Hualing Pampu,
Guangzhou, China

Model :
MINMED GREETMED MEDICAL INSTRUMENTS CO., LTD.
16 P.1 NO.1 Building 1, No.00 Changsan Road,
H. Taid Zone, 31562 Ningbo, Zhejiang Province, China

AXIMED
Feeding Tube, 100 cm
20 pcs

AXIMED
Feeding Tube, 100 cm
20 pcs

AXIMED
Feeding Tube, 100 cm
20 pcs

product of :
AXIONHealthindo

LATEXfree CE0123
STERILE EO

Reg No. HKMKS RI AKL 2005011761

PT. AXION HEALTHINDO
16 P.1 NO.1 Building 1, No.00 Changsan Road,
H. Taid Zone, 31562 Ningbo, Zhejiang Province, China

Disprodksi oleh :
C-4 Jatin Industrial Estate, Hualing Pampu,
Guangzhou, China

Model :
MINMED GREETMED MEDICAL INSTRUMENTS CO., LTD.
16 P.1 NO.1 Building 1, No.00 Changsan Road,
H. Taid Zone, 31562 Ningbo, Zhejiang Province, China

FEATURES :
• Made from non-toxic, non-irritant soft PVC
• Two-eyed tube used for nasogastric feeding of neonates and infants
• Transparent tube for easy observation of fluids
• Sterile, single use only
• Do not use if package is damage or opened

CAUTION :
• Sterile, single use only
• Do not use if package is damage or opened

product of :
AXIONHealthindo

REF :
OD :
type :
size :

LATEXfree CE0123
STERILE EO

Reg No. HKMKS RI AKL 2005011761

PT. AXION HEALTHINDO
16 P.1 NO.1 Building 1, No.00 Changsan Road,
H. Taid Zone, 31562 Ningbo, Zhejiang Province, China

Disprodksi oleh :
C-4 Jatin Industrial Estate, Hualing Pampu,
Guangzhou, China

Model :
MINMED GREETMED MEDICAL INSTRUMENTS CO., LTD.
16 P.1 NO.1 Building 1, No.00 Changsan Road,
H. Taid Zone, 31562 Ningbo, Zhejiang Province, China

AXIMED
Feeding Tube, 100 cm
20 pcs

product of :
AXIONHealthindo

LATEXfree CE0123
STERILE EO

Reg No. HKMKS RI AKL 2005011761

PT. AXION HEALTHINDO
16 P.1 NO.1 Building 1, No.00 Changsan Road,
H. Taid Zone, 31562 Ningbo, Zhejiang Province, China

Disprodksi oleh :
C-4 Jatin Industrial Estate, Hualing Pampu,
Guangzhou, China

Model :
MINMED GREETMED MEDICAL INSTRUMENTS CO., LTD.
16 P.1 NO.1 Building 1, No.00 Changsan Road,
H. Taid Zone, 31562 Ningbo, Zhejiang Province, China



69*53*44cm

 AXIMED FEEDING TUBE, 100 CM CE0123  STERILE EO  product of : AXIONHealthindo	LOT: MFG.: EXP.: QTY.: G.W.: KGS N.W.: KGS MEAS: X X CM	 AXIMED Feeding Tube, 100 cm Dilmpor dan dipasarkan oleh : PT. BIO AXION HEALTHINDO Komplek Golden Plaza Fatmawati Blok G 39-42 Jl. RS Fatmawati No.15, Jakarta Selatan Diproduksi oleh : WELL LEAD MEDICAL CO., LTD. C-4 Jinhu Industrial Estate, Hualong Panyu, Guangzhou, China Melalui : NINGBO GREETMED MEDICAL INSTRUMENTS CO.,LTD. 16 F-1 NO.1 Building 1, No.98 Chuangyuan Road, Hi-Tech Zone, 315042 Ningbo, Zhejiang Province, China Reg No. KEMKES RI AKL 20805911761 LATEXfree CE0123 product of : AXIONHealthindo	Dilmpor dan dipasarkan oleh : PT. BIO AXION HEALTHINDO Komplek Golden Plaza Fatmawati Blok G 39-42 Jl. RS Fatmawati No.15, Jakarta Selatan Diproduksi oleh : WELL LEAD MEDICAL CO., LTD. C-4 Jinhu Industrial Estate, Hualong Panyu, Guangzhou, China Melalui : NINGBO GREETMED MEDICAL INSTRUMENTS CO.,LTD. 16 F-1 NO.1 Building 1, No.98 Chuangyuan Road, Hi-Tech Zone, 315042 Ningbo, Zhejiang Province, China Reg No. KEMKES RI AKL 20805911761
---	--	--	--



fo:d

find

fo:d

fo:d

fo:d

fo:d

bio:

id

fo:d

fo:d

fo:d

fo:d

bioRxiv preprint doi: <https://doi.org/10.1101/000000>; this version posted January 1, 2016. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted bioRxiv a license to display the preprint in perpetuity. It is made available under aCC-BY-NC-ND 4.0 International license.

•

fo:d

fo:d

AXION Healthindo



0123

ID	OD	type	size
REF	LOT		

FEATURES :

- Made from non-toxic, non-irritant soft PVC
- Two-way tube used for nasogastric feeding of neonates and infants as well as diagnostic and therapeutic aspiration
- Transparent tube for easy observation of fluids

CAUTION :

- Sterile, single use only
- Do not use if package is damaged or opened

■ Do not use if package is damaged or opened

•

1106

8 of 8

111

23

9.

product of :
AXIONHealthindo

LATEXfree 0123

②

STERILE	EO
---------	----

 ⚠

ID :	REF :
OD :	 :
type :	 :
size :	 :

LATEXfree CE0123
   

REF :
 LOT :
 :
 :
 product of :
AXIONHealthindo

Ditimpor dan dipasarkan oleh :
PT. BIO AXION HEALTHINDO
Komplek Golden Plaza Fatmawati Blok G 39-42
Jl. RS Fatmawati No.15, Jakarta Selatan

Diproduksi oleh :
WELL LEAD MEDICAL CO., LTD.
C-4 Jinhu Industrial Estate, Hualong Panyu,
Guangzhou, China
Melaka :

NINGBO GREETMED MEDICAL INSTRUMENTS CO.,LTD.
16 F-1 NO.1 Building 1, No.98 Chuangyuan Road,
Hi-Tech Zone, 315042 Ningbo, Zhejiang Province, China

Reg No. KEMKES RI AKL 20805911761

 **AXIMED**
Feeding Tube, 40 cm
20 pcs



AXIMED
Feeding Tube, 40 cm



69*53*44cm

 FEEDING TUBE, 40 CM CE0123  STERILE EO  product of : AXIONHealthindo	LOT: MFG.: EXP.: QTY.: G.W.: KGS N.W.: KGS MEAS: X X CM	 AXIMED Feeding Tube, 40 cm Dilmpor dan dipasarkan oleh : PT. BIO AXION HEALTHINDO Komplek Golden Plaza Fatmawati Blok G 39-42 Jl. RS Fatmawati No.15, Jakarta Selatan Diproduksi oleh : WELL LEAD MEDICAL CO., LTD. C-4 Jinhu Industrial Estate, Hualong Panyu, Guangzhou, China Melalui : NINGBO GREETMED MEDICAL INSTRUMENTS CO.,LTD. 16 F-1 NO.1 Building 1, No.98 Chuangyuan Road, Hi-Tech Zone, 315042 Ningbo, Zhejiang Province, China Reg No. KEMKES RI AKL 20805911761 LATEXfree CE0123 product of :  STERILE EO  AXIONHealthindo	Dilmpor dan dipasarkan oleh : PT. BIO AXION HEALTHINDO Komplek Golden Plaza Fatmawati Blok G 39-42 Jl. RS Fatmawati No.15, Jakarta Selatan Diproduksi oleh : WELL LEAD MEDICAL CO., LTD. C-4 Jinhu Industrial Estate, Hualong Panyu, Guangzhou, China Melalui : NINGBO GREETMED MEDICAL INSTRUMENTS CO.,LTD. 16 F-1 NO.1 Building 1, No.98 Chuangyuan Road, Hi-Tech Zone, 315042 Ningbo, Zhejiang Province, China Reg No. KEMKES RI AKL 20805911761
--	--	--	--



←
PEEL OPEN



Sterile-non pyrogenic. Sterilized by ethylene-oxide gas.
* Do not store at extreme temperatures and humidities.
* Disposable device do not resterilize or reuse.
* Sterile if the package is unopened and undamaged.

LATEXfree CE 0123
⌘ STERILE EO ⚠

product of :
AXIONHealthindo

Dilmpor dan dipasarkan oleh :
PT. BIO AXION HEALTHINDO
Komplek Golden Plaza Fatmawati Blok G 39-42
Jl. RS Fatmawati No.15, Jakarta Selatan

Size

LOT



Diprodukal oleh :
WELL LEAD MEDICAL CO., LTD.
C-4 Jintu Industrial Estate, Hualong Panyu,
Guangzhou, China
Melalui :
NINGBO GREETMED MEDICAL INSTRUMENTS CO.,LTD.
16 F-1 NO.1 Building 1, No.98 Chuangyuan Road,
Hi-Tech Zone, 315042 Ningbo, Zhejiang Province, China



Reg No. KEMKES RI AKL 20805911761

